

Certificate of Validation

Certificate No.	Date
ZFE24072001	2024/7/20

Shenzhen Zafily Technology Co., Ltd. confirms that the products listed in this certificate have been calibrated under the company's ISO9001:2018 quality system, and the product measurement accuracy and performance are in line with the technical indicators declared in the product data sheet.

Reference Instruments

Name	Model	Serial No.
Thermometer	YET-720L	24029071

Environment

Temperature	Humidity
29.3℃	67.3%

All measuring equipment used for calibration is regularly calibrated by a third-party laboratory accredited by CNAS (China National Accreditation Service for Conformity Assessment), and the measurement values of the measuring equipment can be traced back to the national benchmark. CNAS is a full member of the International Laboratory Accreditation Cooperation (www.ilac.org), so the calibration certificate will also be accepted by all members who have signed the Multilateral Mutual Recognition Arrangement (ILAC MRA).

Validation Information

Validation Point	T1	T2	T3
Standard	-30℃	0℃	30℃
Tolerance	± 0.8℃	± 0.5℃	± 0.5℃
Result	PASS	PASS	PASS

Product Information

Model	Serial No.	MFG Date	Expiry Date	Quantity
ZTAG-E	2407131001-2407136020	7/2024	7/2026	5020

The certificate is issued by Shenzhen Zafily Technology Co., Ltd. and is valid for two years from the date of production.

Approved By: